

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 003845  
8:16 21/06/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:49  
PRESSURIZING 3:50  
STERILIZING 4:50  
136.4 C 221kPa 4:50  
136.1 C 222kPa 5:20  
136.4 C 216kPa 5:50  
136.4 C 221kPa 6:20  
136.4 C 224kPa 6:50  
136.1 C 221kPa 7:20  
136.1 C 221kPa 7:50  
136.3 C 223kPa 8:20  
136.3 C 223kPa 8:20  
VENTING 8:20  
AIR DRYING 9:55  
CYCLE COMPLETE 9:59

-----  
Digital Signature #  
556C610EFAD8F749  
-----