
Midmark M11 - v1.0.5
Total Cycles: 2612

Name: RITTER2
Sterilizer ID: V2313228
07/05/2023 03:06 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 198.1 F 0.0
2:00 191.6 F 4.0
4:00 217.1 F 9.3
6:00 218.6 F 7.5
8:00 232.6 F 10.8
10:00 233.1 F 8.5
12:00 239.5 F 10.4
14:00 247.4 F 14.4
16:00 259.9 F 21.3
18:00 271.0 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.0
1:00 272.5 F 29.7
2:00 271.8 F 28.9
3:00 272.1 F 29.8
4:00 272.3 F 29.4

Min 271.3 F 28.8
Max 272.8 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.5 F 0.0
5:00 223.1 F 0.0
10:00 226.3 F 0.0
15:00 224.5 F 0.0
20:00 221.6 F 0.0
25:00 222.7 F 0.0
30:00 221.9 F 0.0
Drying Complete

FILLING: 2:10
HEATING: 18:44
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:58:08
CYCLE COMPLETE

07/05/2023 04:04 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____