

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007

Unit # : 000

WATER QUALITY

1.3uS / 0.8ppm

CYCLE NUMBER 000810

14:32 17/11/2021

RUBBER AND PLASTICS

121 C for 15 min.

CYCLE START 0:00

CONDITIONING 1:44

PRESSURIZING 3:09

STERILIZING 3:31

123.0 C 122kPa 3:31

123.4 C 124kPa 4:01

123.1 C 122kPa 4:31

123.3 C 123kPa 5:01

123.4 C 124kPa 5:31

123.2 C 124kPa 6:01

123.4 C 122kPa 6:31

123.4 C 125kPa 7:01

123.4 C 124kPa 7:31

123.4 C 125kPa 8:01

123.4 C 124kPa 8:31

123.2 C 123kPa 9:01

123.4 C 123kPa 9:31

123.2 C 124kPa 10:01

123.4 C 124kPa 10:31

123.4 C 124kPa 11:01

123.2 C 124kPa 11:31

123.2 C 123kPa 12:01

123.4 C 122kPa 12:31

123.3 C 123kPa 13:01

123.4 C 124kPa 13:31

123.4 C 124kPa 14:01

123.4 C 124kPa 14:31

123.4 C 125kPa 15:01

123.4 C 125kPa 15:31

123.4 C 122kPa 16:01

123.4 C 124kPa 16:31

123.4 C 125kPa 17:01

123.3 C 124kPa 17:31

123.4 C 125kPa 18:01

123.4 C 123kPa 18:31

123.4 C 123kPa 18:31

VENTING 18:31  
AIR DRYING 19:12  
CYCLE COMPLETE 29:13

-----  
Digital Signature #  
27F5384A723EFC87  
-----