
Midmark M11 - v1.0.0
Total Cycles: 3033

Name: RITTER 1
Sterilizer ID: V2178198
03/24/2023 12:32 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 181.4 F 0.0
2:00 175.2 F 3.5
4:00 191.5 F 7.8
6:00 217.0 F 13.3
8:00 226.5 F 7.8
10:00 239.2 F 12.9
12:00 237.4 F 10.1
14:00 237.8 F 8.2
16:00 249.3 F 15.0
18:00 261.2 F 21.7
20:00 272.2 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.4
1:00 272.4 F 28.9
2:00 271.3 F 28.5
3:00 272.5 F 29.0
4:00 271.5 F 28.3

Min 271.2 F 28.3
Max 272.6 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.2 F 0.0
5:00 226.0 F 0.0
10:00 225.5 F 0.0
15:00 230.8 F 0.0
20:00 227.5 F 0.0
25:00 224.9 F 0.0
30:00 204.4 F 0.0
Drying Complete

FILLING: 1:26
HEATING: 20:29
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 00:59:12
CYCLE COMPLETE

03/24/2023 01:31 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____