
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001897
10:02 26/07/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:00
PRESSURIZING 2:06
STERILIZING 3:06
136.0 C 220kPa 3:06
136.4 C 225kPa 3:36
136.4 C 225kPa 4:06
136.2 C 223kPa 4:36
136.3 C 223kPa 5:06
136.4 C 224kPa 5:36
136.4 C 223kPa 6:06
136.4 C 225kPa 6:36
136.4 C 225kPa 6:36
VENTING 6:36
AIR DRYING 7:27
CYCLE COMPLETE 21:35

Digital Signature #
FA957C265742BA54
