
Midmark M11 - v1.0.0
Total Cycles: 1091

Name: RITTER 1
Sterilizer ID: V2178198
08/18/2021 08:48 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 175.3 F 0.0
2:00 166.6 F 1.6
4:00 171.1 F 4.9
6:00 195.5 F 9.1
8:00 222.3 F 9.7
10:00 219.1 F 7.4
12:00 234.5 F 12.3
14:00 234.9 F 9.2
16:00 236.7 F 9.2
18:00 247.1 F 13.5
20:00 258.7 F 19.8
22:00 268.8 F 26.6

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.8
1:00 272.0 F 28.7
2:00 272.5 F 29.0
3:00 272.3 F 29.2
4:00 272.3 F 29.3

Min 271.4 F 28.7
Max 272.9 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.8 F 0.0
5:00 236.4 F 0.0
10:00 225.3 F 0.0
15:00 220.3 F 0.0
20:00 220.6 F 0.0
25:00 227.3 F 0.0
30:00 229.8 F 0.0
Drying Complete

FILLING: 1:09
HEATING: 23:50
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 01:02:06
CYCLE COMPLETE

08/18/2021 09:50 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____