
Midmark M11 - v1.0.5
Total Cycles: 1717

Name: RITTER2
Sterilizer ID: V2313228
11/14/2022 01:24 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 176.0 F 0.0
2:00 169.1 F 0.2
4:00 192.0 F 2.0
6:00 214.2 F 7.0
8:00 227.2 F 12.2
10:00 231.8 F 9.0
12:00 239.3 F 9.9
14:00 245.3 F 13.2
16:00 245.4 F 13.1
18:00 258.3 F 20.1
20:00 269.7 F 27.5

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.7
1:00 272.5 F 29.3
2:00 271.3 F 28.8
3:00 272.6 F 29.3
4:00 271.5 F 28.6

Min 271.3 F 28.6
Max 272.9 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.6 F 0.0
5:00 220.7 F 0.0
10:00 242.6 F 0.0
15:00 231.2 F 0.0
20:00 228.9 F 0.0
25:00 226.4 F 0.0
30:00 223.6 F 0.0
Drying Complete

FILLING: 1:10
HEATING: 20:56
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:59:15
CYCLE COMPLETE

11/14/2022 02:23 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____