
Midmark M11 - v1.0.0
Total Cycles: 4486

Name: RITTER 1
Sterilizer ID: V2178198
04/24/2024 12:29 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 166.4 F 0.0
2:00 162.9 F 2.6
4:00 166.8 F 6.1
6:00 201.4 F 9.9
8:00 216.0 F 5.9
10:00 227.5 F 8.0
12:00 233.0 F 12.6
14:00 234.5 F 9.2
16:00 241.0 F 10.1
18:00 245.0 F 13.1
20:00 256.7 F 19.2
22:00 267.0 F 25.5

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.6
1:00 271.9 F 29.3
2:00 272.1 F 28.9
3:00 271.4 F 28.9
4:00 272.1 F 29.2

Min 271.1 F 28.5
Max 272.9 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.6 F 0.0
5:00 226.8 F 0.0
10:00 233.4 F 0.0
15:00 226.6 F 0.0
20:00 227.5 F 0.0
25:00 228.6 F 0.0
30:00 230.6 F 0.0
Drying Complete

FILLING: 1:04
HEATING: 23:26
STERILIZING: 4:00
VENTING: 3:42
DRYING: 30:00
TOTAL CYCLE: 01:02:12
CYCLE COMPLETE

04/24/2024 01:31 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____