
Midmark M11 - v1.0.0
Total Cycles: 5043

Name: RITTER 1
Sterilizer ID: V2178198
11/01/2024 08:37 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.8 F 0.0
2:00 78.6 F 0.0
4:00 85.9 F 0.0
6:00 118.7 F 0.1
8:00 143.4 F 0.3
10:00 172.8 F 1.4
12:00 192.6 F 4.5
14:00 207.1 F 7.9
16:00 219.5 F 11.4
18:00 219.2 F 4.8
20:00 232.2 F 9.5
22:00 237.5 F 10.7
24:00 241.0 F 11.3
26:00 236.9 F 9.2
28:00 250.9 F 15.7
30:00 262.2 F 22.1
32:00 272.7 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.2
1:00 272.7 F 29.0
2:00 271.9 F 28.3
3:00 271.8 F 28.6
4:00 272.7 F 29.0

Min 271.4 F 28.1
Max 272.8 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 174.1 F 0.0
5:00 209.8 F 0.0
10:00 217.0 F 0.0
15:00 223.8 F 0.0
20:00 226.2 F 0.0
25:00 231.9 F 0.0
30:00 230.8 F 0.0

Drying Complete

FILLING: 0:43
HEATING: 32:25
STERILIZING: 4:00
VENTING: 3:29
DRYING: 30:00
TOTAL CYCLE: 01:10:37
CYCLE COMPLETE

11/01/2024 09:47 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____