

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1200

Name: RITTER2  
Sterilizer ID: V2313228  
06/20/2022 09:00 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 123.9 F 0.0  
2:00 126.5 F 1.2  
4:00 133.9 F 3.5  
6:00 160.5 F 7.1  
8:00 187.5 F 11.4  
10:00 215.4 F 4.7  
12:00 222.7 F 9.6  
14:00 234.5 F 9.2  
16:00 239.2 F 11.3  
18:00 237.1 F 9.3  
20:00 250.1 F 15.6  
22:00 261.2 F 22.1  
24:00 271.5 F 28.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.3  
1:00 272.9 F 29.6  
2:00 271.5 F 29.6  
3:00 272.3 F 29.6  
4:00 271.6 F 29.7

Min 271.3 F 29.2  
Max 273.0 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 186.6 F 0.0  
5:00 216.6 F 0.0  
10:00 223.5 F 0.0  
15:00 224.8 F 0.0  
20:00 222.4 F 0.0  
25:00 224.2 F 0.0  
30:00 225.1 F 0.0  
Drying Complete

FILLING: 0:56  
HEATING: 25:29  
STERILIZING: 4:00  
VENTING: 3:06  
DRYING: 30:00  
TOTAL CYCLE: 01:03:31  
CYCLE COMPLETE

06/20/2022 10:03 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_