
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000653
9:57 13/09/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:39
PRESSURIZING 2:47
STERILIZING 4:14
136.5 C 233kPa 4:14
136.4 C 233kPa 4:44
136.3 C 231kPa 5:14
136.2 C 230kPa 5:44
136.2 C 231kPa 6:14
136.2 C 229kPa 6:44
136.4 C 233kPa 7:14
136.4 C 233kPa 7:44
136.4 C 233kPa 7:44
VENTING 7:44
AIR DRYING 8:40
CYCLE COMPLETE 8:49

Digital Signature #
13550537E6A11D69
