
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 002169
7:26 04/08/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:20
PRESSURIZING 2:22
STERILIZING 3:20
136.4 C 223kPa 3:20
136.1 C 218kPa 3:50
136.4 C 223kPa 4:20
136.5 C 222kPa 4:50
136.3 C 223kPa 5:20
136.4 C 223kPa 5:50
136.3 C 223kPa 6:20
136.4 C 224kPa 6:50
136.4 C 224kPa 6:50
VENTING 6:50
AIR DRYING 7:40
CYCLE COMPLETE 7:41

Digital Signature #
6D71B3A52E1ADDBF
