
Midmark M11 - v1.0.0
Total Cycles: 3123

Name: RITTER 1
Sterilizer ID: V2178198
04/18/2023 01:05 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 169.6 F 0.0
2:00 167.0 F 0.0
4:00 190.5 F 3.1
6:00 208.6 F 8.2
8:00 222.3 F 13.2
10:00 229.2 F 8.3
12:00 241.3 F 13.5
14:00 240.7 F 11.4
16:00 238.7 F 10.0
18:00 252.6 F 16.8
20:00 264.4 F 23.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.4 F 29.0
2:00 271.5 F 28.5
3:00 272.5 F 29.1
4:00 271.6 F 28.4

Min 271.4 F 28.3
Max 272.8 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.3 F 0.0
5:00 217.2 F 0.0
10:00 230.9 F 0.0
15:00 229.7 F 0.0
20:00 228.9 F 0.0
25:00 229.0 F 0.0
30:00 227.3 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 21:55
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 01:00:04
CYCLE COMPLETE

04/18/2023 02:05 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____