

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3389

Name: RITTER 1  
Sterilizer ID: V2178198  
06/14/2023 12:46 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 203.9 F 0.0  
2:00 194.2 F 2.5  
4:00 211.7 F 7.9  
6:00 227.4 F 13.5  
8:00 233.0 F 9.5  
10:00 237.9 F 8.9  
12:00 245.3 F 12.8  
14:00 245.0 F 12.9  
16:00 258.1 F 19.8  
18:00 269.5 F 27.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.4  
1:00 272.5 F 28.9  
2:00 271.7 F 28.6  
3:00 272.4 F 28.9  
4:00 271.4 F 28.3  
  
Min 271.4 F 28.2  
Max 272.8 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.7 F 0.0  
5:00 225.3 F 0.0  
10:00 238.7 F 0.0  
15:00 231.8 F 0.0  
20:00 225.2 F 0.0  
25:00 224.3 F 0.0  
30:00 227.7 F 0.0  
Drying Complete

FILLING: 2:01  
HEATING: 18:56  
STERILIZING: 4:00  
VENTING: 3:22  
DRYING: 30:00  
TOTAL CYCLE: 00:58:19  
CYCLE COMPLETE

06/14/2023 01:44 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_