
Midmark M11 - v1.0.5
Total Cycles: 4264

Name: RITTER2
Sterilizer ID: V2313228
01/14/2025 05:11 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 131.2 F 0.0
2:00 130.3 F 0.0
4:00 144.7 F 0.3
6:00 176.4 F 1.1
8:00 197.9 F 4.9
10:00 218.3 F 9.3
12:00 226.9 F 10.1
14:00 233.9 F 10.3
16:00 235.0 F 8.6
18:00 245.0 F 13.0
20:00 244.2 F 12.7
22:00 256.6 F 19.3
24:00 267.8 F 26.2

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.8
1:00 272.2 F 29.2
2:00 271.8 F 29.3
3:00 272.0 F 29.0
4:00 272.0 F 29.6

Min 271.3 F 28.7
Max 272.6 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.0 F 0.0
5:00 214.1 F 0.0
10:00 187.2 F 0.0
15:00 185.7 F 0.0
20:00 184.1 F 0.0
25:00 183.0 F 0.0
30:00 182.3 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 25:19
STERILIZING: 4:00
VENTING: 3:32
DRYING: 30:00
TOTAL CYCLE: 01:03:45
CYCLE COMPLETE

01/14/2025 06:14 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____