
Midmark M11 - v1.0.5
Total Cycles: 459

Name: RITTER2
Sterilizer ID: V2313228
10/19/2021 09:22 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 143.2 F 0.0
2:00 144.7 F 1.4
4:00 149.1 F 4.5
6:00 174.8 F 8.6
8:00 195.9 F 13.3
10:00 215.3 F 7.3
12:00 229.7 F 12.6
14:00 233.5 F 9.6
16:00 236.5 F 9.5
18:00 246.6 F 14.2
20:00 258.5 F 20.7
22:00 269.2 F 27.6

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.1
1:00 272.4 F 30.0
2:00 271.9 F 29.5
3:00 271.1 F 29.0
4:00 271.6 F 29.8

Min 271.0 F 29.0
Max 272.8 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.1 F 0.0
5:00 205.1 F 0.0
10:00 205.5 F 0.0
15:00 203.9 F 0.0
20:00 194.8 F 0.0
25:00 189.2 F 0.0
30:00 193.5 F 0.0
Drying Complete

FILLING: 1:13
HEATING: 23:06
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 01:01:27
CYCLE COMPLETE

10/19/2021 10:23 AM
ACCEPTED: X
REJECTED:
Operator: U01