
Midmark M11 - v1.0.5
Total Cycles: 4452

Name: RITTER2
Sterilizer ID: V2313228
03/25/2025 11:05 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 169.4 F 0.0
2:00 164.7 F 1.5
4:00 164.1 F 4.7
6:00 192.3 F 8.6
8:00 200.9 F 13.0
10:00 221.5 F 7.4
12:00 235.7 F 11.9
14:00 232.5 F 8.9
16:00 241.2 F 11.2
18:00 244.3 F 13.0
20:00 255.2 F 19.0
22:00 265.4 F 25.2
24:00 270.8 F 29.6

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 30.4
1:00 272.1 F 30.7
2:00 272.1 F 30.7
3:00 272.0 F 30.7
4:00 271.8 F 30.6

Min 271.8 F 30.3
Max 272.4 F 30.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.3 F 0.0
5:00 225.3 F 0.0
10:00 220.1 F 0.0
15:00 216.5 F 0.0
20:00 214.2 F 0.0
25:00 214.4 F 0.0
30:00 212.0 F 0.0
Drying Complete

FILLING: 1:22
HEATING: 24:41
STERILIZING: 4:00
VENTING: 3:58
DRYING: 30:00
TOTAL CYCLE: 01:04:01
CYCLE COMPLETE

03/25/2025 12:08 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____