
Midmark M11 - v1.0.5
Total Cycles: 2764

Name: RITTER2
Sterilizer ID: V2313228
08/14/2023 05:59 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 152.4 F 0.0
2:00 153.0 F 0.0
4:00 177.6 F 1.2
6:00 198.0 F 5.2
8:00 214.0 F 9.4
10:00 225.4 F 9.6
12:00 233.7 F 10.0
14:00 235.5 F 9.2
16:00 243.8 F 12.4
18:00 245.5 F 13.4
20:00 257.5 F 19.9
22:00 268.2 F 26.6

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.9
1:00 272.2 F 29.1
2:00 272.5 F 29.6
3:00 271.6 F 28.8
4:00 272.4 F 29.3

Min 271.5 F 28.7
Max 272.6 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 179.3 F 0.0
5:00 228.4 F 0.0
10:00 228.0 F 0.0
15:00 227.4 F 0.0
20:00 224.3 F 0.0
25:00 222.0 F 0.0
30:00 219.3 F 0.0
Drying Complete

FILLING: 1:27
HEATING: 23:13
STERILIZING: 4:00
VENTING: 3:29
DRYING: 30:00
TOTAL CYCLE: 01:02:09
CYCLE COMPLETE

08/14/2023 07:01 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____