
Midmark M11 - v1.0.0
Total Cycles: 1754

Name: RITTER 1
Sterilizer ID: V2178198
03/28/2022 12:24 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 175.7 F 0.0
2:00 172.1 F 3.1
4:00 183.5 F 6.8
6:00 219.5 F 11.6
8:00 223.0 F 6.2
10:00 235.6 F 11.5
12:00 233.2 F 8.3
14:00 242.1 F 11.0
16:00 245.1 F 12.9
18:00 257.6 F 19.6
20:00 268.8 F 26.6

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.4
1:00 272.6 F 29.1
2:00 271.5 F 28.4
3:00 272.0 F 29.1
4:00 272.3 F 28.9

Min 271.2 F 28.3
Max 272.6 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.3 F 0.0
5:00 229.7 F 0.0
10:00 228.2 F 0.0
15:00 233.7 F 0.0
20:00 226.7 F 0.0
25:00 231.2 F 0.0
30:00 227.7 F 0.0
Drying Complete

FILLING: 1:11
HEATING: 21:09
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 00:59:27
CYCLE COMPLETE

03/28/2022 01:23 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____