
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 002394
8:51 18/12/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:04
PRESSURIZING 3:06
STERILIZING 4:14
136.0 C 220kPa 4:14
136.3 C 224kPa 4:44
136.2 C 222kPa 5:14
136.2 C 222kPa 5:44
136.4 C 225kPa 6:14
136.4 C 222kPa 6:44
136.1 C 223kPa 7:14
136.2 C 223kPa 7:44
136.2 C 223kPa 7:44
VENTING 7:44
AIR DRYING 8:34
CYCLE COMPLETE 38:34

Digital Signature #
43ACA865376FCD33
