
Midmark M11 - v1.0.5
Total Cycles: 4405

Name: RITTER2
Sterilizer ID: V2313228
03/06/2025 09:43 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.8 F 0.0
2:00 71.1 F 0.2
4:00 76.8 F 1.1
6:00 102.0 F 3.2
8:00 135.7 F 6.3
10:00 162.8 F 10.1
12:00 196.4 F 6.6
14:00 208.6 F 5.0
16:00 219.4 F 8.5
18:00 230.2 F 12.4
20:00 230.4 F 8.6
22:00 241.8 F 12.9
24:00 241.0 F 11.9
26:00 253.1 F 17.9
28:00 263.7 F 24.3
30:00 272.2 F 29.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.5
1:00 272.7 F 30.1
2:00 271.8 F 29.7
3:00 272.6 F 30.0
4:00 271.8 F 29.8

Min 271.5 F 29.4
Max 272.8 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.4 F 0.0
5:00 209.2 F 0.0
10:00 189.8 F 0.0
15:00 191.2 F 0.0
20:00 196.1 F 0.0
25:00 215.2 F 0.0
30:00 221.3 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 30:10
STERILIZING: 4:00
VENTING: 3:28
DRYING: 30:00
TOTAL CYCLE: 01:08:30
CYCLE COMPLETE

03/06/2025 10:51 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____