

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 000454  
13:11 01/02/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:07  
PRESSURIZING 3:08  
STERILIZING 4:13  
136.0 C 228kPa 4:13  
136.3 C 233kPa 4:43  
136.4 C 233kPa 5:13  
136.2 C 231kPa 5:43  
136.4 C 232kPa 6:13  
136.4 C 231kPa 6:43  
136.4 C 233kPa 7:13  
136.4 C 233kPa 7:43  
136.4 C 233kPa 7:43  
VENTING 7:43  
AIR DRYING 8:38  
CYCLE COMPLETE 8:40  
-----

Digital Signature #  
AF0E30E07DD7BD50  
-----