
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000854
16:36 07/01/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:25
PRESSURIZING 2:33
STERILIZING 3:37
136.0 C 229kPa 3:37
136.2 C 231kPa 4:07
136.4 C 233kPa 4:37
136.1 C 230kPa 5:07
136.3 C 233kPa 5:37
136.4 C 233kPa 6:07
136.4 C 233kPa 6:37
136.2 C 231kPa 7:07
136.2 C 231kPa 7:07
VENTING 7:07
AIR DRYING 7:59
CYCLE COMPLETE 22:59

Digital Signature #
9B56295000A1CCBB
