
Midmark M11 - v1.0.5
Total Cycles: 2059

Name: RITTER2
Sterilizer ID: V2313228
02/15/2023 04:37 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 203.1 F 0.0
2:00 202.3 F 3.9
4:00 223.5 F 9.9
6:00 229.2 F 6.8
8:00 239.9 F 13.3
10:00 241.7 F 12.1
12:00 242.3 F 11.8
14:00 256.7 F 19.4
16:00 269.1 F 27.2

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.9
1:00 271.9 F 28.8
2:00 272.3 F 29.1
3:00 272.8 F 29.5
4:00 272.1 F 29.4

Min 271.6 F 28.7
Max 272.8 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.9 F 0.0
5:00 220.8 F 0.0
10:00 235.3 F 0.0
15:00 235.8 F 0.0
20:00 226.6 F 0.0
25:00 221.4 F 0.0
30:00 220.3 F 0.0

Drying Complete

FILLING: 2:31
HEATING: 16:57
STERILIZING: 4:00
VENTING: 3:04
DRYING: 30:00
TOTAL CYCLE: 00:56:32
CYCLE COMPLETE

02/15/2023 05:33 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____