
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000423
11:21 27/01/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:05
PRESSURIZING 2:12
STERILIZING 3:12
136.0 C 222kPa 3:12
136.2 C 231kPa 3:42
136.1 C 230kPa 4:12
136.4 C 233kPa 4:42
136.2 C 232kPa 5:12
136.1 C 231kPa 5:42
136.1 C 232kPa 6:12
136.4 C 233kPa 6:42
136.4 C 233kPa 6:42
VENTING 6:43
AIR DRYING 7:36
CYCLE COMPLETE 7:51

Digital Signature #
E94EE2B8C324B4E3
