
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000165
8:55 18/06/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:05
PRESSURIZING 3:09
STERILIZING 4:12
136.0 C 220kPa 4:12
135.7 C 219kPa 4:43
136.4 C 221kPa 5:13
136.4 C 220kPa 5:43
136.2 C 223kPa 6:13
136.3 C 223kPa 6:43
136.3 C 224kPa 7:13
136.4 C 224kPa 7:43
136.4 C 224kPa 7:43
VENTING 7:43
AIR DRYING 8:35
CYCLE COMPLETE 8:48

Digital Signature #
5011253383A7478C
