
Midmark M11 - v1.0.0
Total Cycles: 1082

Name: RITTER 1
Sterilizer ID: V2178198
08/09/2021 03:37 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 141.9 F 0.0
2:00 143.6 F 1.3
4:00 164.6 F 3.5
6:00 182.9 F 6.8
8:00 202.2 F 10.9
10:00 220.6 F 4.5
12:00 232.0 F 8.8
14:00 242.3 F 13.5
16:00 240.5 F 10.9
18:00 236.7 F 8.9
20:00 250.3 F 15.2
22:00 261.2 F 21.4
24:00 271.4 F 28.0

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.1
1:00 272.7 F 28.9
2:00 271.9 F 28.3
3:00 271.3 F 28.4
4:00 272.6 F 28.9

Min 271.0 F 28.0
Max 272.8 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 173.9 F 0.0
5:00 195.5 F 0.0
10:00 194.1 F 0.0
15:00 184.9 F 0.0
20:00 179.4 F 0.0
25:00 177.7 F 0.0
30:00 169.7 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 24:43
STERILIZING: 4:00
VENTING: 3:26
DRYING: 30:00
TOTAL CYCLE: 01:03:07
CYCLE COMPLETE

08/09/2021 04:40 PM
ACCEPTED: X
REJECTED:
Operator: U01