
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001536
16:50 23/10/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:30
PRESSURIZING 2:31
STERILIZING 3:35
136.0 C 223kPa 3:35
136.4 C 227kPa 4:05
136.4 C 224kPa 4:35
136.4 C 225kPa 5:05
136.4 C 226kPa 5:35
136.4 C 227kPa 6:05
136.4 C 226kPa 6:35
136.4 C 225kPa 7:05
136.4 C 225kPa 7:05
VENTING 7:05
AIR DRYING 8:00
CYCLE COMPLETE 8:12

Digital Signature #
8C6F99190D86C8AD
