

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2766

Name: RITTER2  
Sterilizer ID: V2313228  
08/15/2023 08:36 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 201.9 F 0.0  
2:00 196.7 F 0.0  
4:00 208.1 F 0.0  
6:00 211.9 F 0.0  
8:00 222.7 F 3.6  
10:00 235.6 F 8.5  
12:00 246.3 F 12.8  
14:00 244.7 F 12.7  
16:00 243.5 F 12.1  
18:00 243.0 F 11.9  
20:00 256.5 F 18.9  
22:00 268.2 F 26.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 28.5  
1:00 272.2 F 28.6  
2:00 272.7 F 29.0  
3:00 272.6 F 29.3  
4:00 271.8 F 28.5

Min 271.7 F 28.3  
Max 272.9 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 183.3 F 0.0  
5:00 208.9 F 0.0  
10:00 203.0 F 0.0  
15:00 202.7 F 0.0  
20:00 196.7 F 0.0  
25:00 193.4 F 0.0  
30:00 193.5 F 0.0  
Drying Complete

FILLING: 0:50  
HEATING: 23:11  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 01:01:11  
CYCLE COMPLETE

08/15/2023 09:37 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_