
Midmark M11 - v1.0.5
Total Cycles: 2066

Name: RITTER2
Sterilizer ID: V2313228
02/17/2023 11:48 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 74.4 F 0.0
2:00 92.5 F 0.0
4:00 87.2 F 0.0
6:00 110.2 F 0.0
8:00 144.7 F 1.5
10:00 176.3 F 3.7
12:00 201.0 F 7.0
14:00 221.7 F 11.0
16:00 220.2 F 6.0
18:00 233.0 F 9.9
20:00 235.3 F 9.6
22:00 242.9 F 12.8
24:00 241.8 F 11.7
26:00 254.5 F 18.2
28:00 266.1 F 25.2

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 29.0
1:00 272.8 F 29.7
2:00 271.7 F 29.3
3:00 272.4 F 29.3
4:00 272.1 F 29.7

Min 271.3 F 28.9
Max 272.8 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.5 F 0.0
5:00 236.0 F 0.0
10:00 232.5 F 0.0
15:00 224.6 F 0.0
20:00 224.1 F 0.0
25:00 229.7 F 0.0
30:00 223.4 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 29:31
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 01:07:39
CYCLE COMPLETE

02/17/2023 12:55 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____