

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
2.5uS / 1.6ppm  
CYCLE NUMBER 001018  
13:29 25/03/2022  
WRAPPED  
135 C for 10 min.  
CYCLE START 0:00  
CONDITIONING 0:59  
PRESSURIZING 2:47  
STERILIZING 3:41  
136.4 C 231kPa 3:41  
136.2 C 231kPa 4:11  
136.4 C 232kPa 4:41  
136.2 C 230kPa 5:11  
136.3 C 231kPa 5:41  
136.4 C 233kPa 6:11  
136.2 C 231kPa 6:41  
136.4 C 232kPa 7:11  
136.2 C 228kPa 7:41  
136.4 C 228kPa 8:11  
136.4 C 231kPa 8:41  
136.4 C 233kPa 9:11  
136.3 C 232kPa 9:41  
CYCLE ABORTED 10:01  
STOP BUTTON PRESSED  
NOT STERILE  
END TIME 10:01  
-----

Digital Signature #  
28ADE9C8580B5EEE  
-----