

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 002370  
14:03 10/12/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:19  
PRESSURIZING 3:21  
STERILIZING 4:22  
136.4 C 220kPa 4:22  
136.4 C 225kPa 4:52  
136.4 C 225kPa 5:22  
136.4 C 224kPa 5:52  
136.0 C 222kPa 6:22  
136.3 C 224kPa 6:52  
136.1 C 221kPa 7:22  
136.4 C 222kPa 7:52  
136.4 C 222kPa 7:52  
VENTING 7:52  
AIR DRYING 8:42  
CYCLE COMPLETE 38:42

-----  
Digital Signature #  
F7B0C5B17B437C88  
-----