
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002001
10:39 10/09/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:06
PRESSURIZING 2:10
STERILIZING 3:06
136.0 C 220kPa 3:06
136.2 C 223kPa 3:37
136.4 C 225kPa 4:07
136.4 C 226kPa 4:37
136.4 C 225kPa 5:07
136.2 C 222kPa 5:37
136.4 C 226kPa 6:07
136.3 C 224kPa 6:37
136.3 C 224kPa 6:37
VENTING 6:37
AIR DRYING 7:30
CYCLE COMPLETE 8:00

Digital Signature #
30A19B0483A3C138
