

-----  
Midmark M11 - v1.0.5  
Total Cycles: 362

Name: RITTER2  
Sterilizer ID: V2313228  
09/14/2021 12:38 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 189.6 F 0.0  
2:00 188.4 F 3.9  
4:00 204.9 F 10.1  
6:00 221.6 F 6.5  
8:00 233.7 F 13.1  
10:00 237.8 F 11.8  
12:00 240.2 F 11.4  
14:00 254.5 F 18.6  
16:00 267.0 F 25.9

STERILIZING  
mm:ss Degrees PSI  
0:00 272.1 F 29.8  
1:00 272.4 F 30.2  
2:00 272.8 F 30.4  
3:00 272.7 F 30.3  
4:00 272.7 F 30.3  
  
Min 272.1 F 29.7  
Max 273.4 F 30.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 191.9 F 0.0  
5:00 236.1 F 0.0  
10:00 226.0 F 0.0  
15:00 216.7 F 0.0  
Dry Cycle Aborted

FILLING: 2:25  
HEATING: 17:57  
STERILIZING: 4:00  
VENTING: 3:00  
DRYING: 19:04  
TOTAL CYCLE: 00:46:26

09/14/2021 01:29 PM  
ACCEPTED: X  
REJECTED:  
Operator: U01