
Midmark M11 - v1.0.5
Total Cycles: 695

Name: RITTER2
Sterilizer ID: V2313228
01/13/2022 11:54 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 200.4 F 0.0
2:00 191.2 F 4.2
4:00 203.9 F 9.4
6:00 222.7 F 7.2
8:00 236.6 F 10.8
10:00 234.0 F 8.7
12:00 239.3 F 10.2
14:00 248.8 F 15.0
16:00 261.8 F 22.3
18:00 272.7 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.8
1:00 271.7 F 28.8
2:00 272.2 F 29.2
3:00 271.8 F 29.4
4:00 271.6 F 28.8

Min 271.2 F 28.6
Max 272.4 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.5 F 0.0
5:00 236.2 F 0.0
10:00 220.0 F 0.0
15:00 232.0 F 0.0
20:00 221.1 F 0.0
25:00 223.2 F 0.0
30:00 216.2 F 0.0
Drying Complete

FILLING: 2:35
HEATING: 18:14
STERILIZING: 4:00
VENTING: 3:05
DRYING: 30:00
TOTAL CYCLE: 00:57:54
CYCLE COMPLETE

01/13/2022 01:13 PM
ACCEPTED: X
REJECTED:
Operator: U02