
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000410
13:04 26/01/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:58
PRESSURIZING 3:04
STERILIZING 4:06
136.0 C 230kPa 4:06
136.1 C 230kPa 4:36
136.3 C 232kPa 5:06
136.2 C 231kPa 5:36
136.2 C 231kPa 6:06
136.1 C 230kPa 6:36
136.2 C 232kPa 7:06
136.4 C 232kPa 7:36
136.4 C 232kPa 7:36
VENTING 7:36
AIR DRYING 8:31
CYCLE COMPLETE 8:32

Digital Signature #
B9AC12A63C4126B9
