

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3893

Name: RITTER 1  
Sterilizer ID: V2178198  
10/12/2023 09:08 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 174.3 F 0.0  
2:00 171.9 F 2.4  
4:00 186.0 F 6.5  
6:00 228.5 F 11.9  
8:00 222.5 F 7.1  
10:00 238.4 F 12.6  
12:00 236.9 F 9.6  
14:00 237.1 F 8.7  
16:00 248.5 F 14.5  
18:00 260.9 F 21.3  
20:00 272.1 F 28.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.2  
1:00 272.8 F 29.1  
2:00 272.6 F 29.2  
3:00 272.1 F 29.1  
4:00 271.5 F 28.7

Min 270.8 F 28.1  
Max 273.0 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 182.5 F 0.0  
5:00 243.9 F 0.0  
10:00 219.8 F 0.0  
15:00 228.6 F 0.0  
20:00 216.6 F 0.0  
25:00 225.8 F 0.0  
30:00 219.3 F 0.0  
Drying Complete

FILLING: 1:39  
HEATING: 20:24  
STERILIZING: 4:00  
VENTING: 3:20  
DRYING: 30:00  
TOTAL CYCLE: 00:59:24  
CYCLE COMPLETE

10/12/2023 10:07 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_