
Midmark M11 - v1.0.5
Total Cycles: 3885

Name: RITTER2
Sterilizer ID: V2313228
08/22/2024 03:30 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 134.1 F 0.0
2:00 134.0 F 1.4
4:00 138.3 F 3.9
6:00 165.7 F 7.6
8:00 186.1 F 11.8
10:00 211.8 F 5.2
12:00 224.0 F 9.6
14:00 230.9 F 10.2
16:00 237.1 F 11.3
18:00 236.1 F 9.5
20:00 249.1 F 15.6
22:00 260.1 F 21.8
24:00 270.5 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.9
1:00 272.1 F 29.8
2:00 272.9 F 30.1
3:00 272.7 F 30.4
4:00 271.5 F 30.0

Min 271.5 F 29.8
Max 273.0 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.7 F 0.0
5:00 217.2 F 0.0
10:00 213.0 F 0.0
15:00 212.7 F 0.0
20:00 212.3 F 0.0
25:00 225.2 F 0.0
30:00 219.5 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 25:33
STERILIZING: 4:00
VENTING: 3:45
DRYING: 30:00
TOTAL CYCLE: 01:04:20
CYCLE COMPLETE

08/22/2024 04:34 PM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____