
Midmark M11 - v1.0.5
Total Cycles: 2085

Name: RITTER2
Sterilizer ID: V2313228
02/23/2023 12:13 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 195.9 F 0.0
2:00 189.0 F 3.2
4:00 202.4 F 8.3
6:00 220.9 F 10.0
8:00 230.9 F 10.3
10:00 232.7 F 8.1
12:00 240.1 F 10.5
14:00 248.5 F 14.8
16:00 261.3 F 22.1
18:00 272.8 F 29.6

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.8
1:00 272.5 F 29.5
2:00 271.8 F 28.8
3:00 271.6 F 29.2
4:00 272.7 F 29.5

Min 271.2 F 28.6
Max 272.8 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.1 F 0.0
5:00 233.4 F 0.0
10:00 232.2 F 0.0
15:00 226.8 F 0.0
20:00 221.4 F 0.0
25:00 220.1 F 0.0
30:00 220.0 F 0.0
Drying Complete

FILLING: 1:51
HEATING: 18:24
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:57:30
CYCLE COMPLETE

02/23/2023 01:10 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____