
Midmark M11 - v1.0.0
Total Cycles: 715

Name: MIDMARK
Sterilizer ID: V2178198
03/01/2021 02:40 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 150.4 F 0.0
2:00 151.4 F 2.1
4:00 158.9 F 5.8
6:00 231.7 F 11.1
8:00 224.5 F 6.4
10:00 232.7 F 12.2
12:00 237.3 F 9.3
14:00 236.4 F 8.6
16:00 248.6 F 14.4
18:00 261.4 F 21.6
20:00 272.7 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 28.8
1:00 272.5 F 29.2
2:00 272.6 F 28.9
3:00 272.1 F 28.8
4:00 272.1 F 29.2

Min 271.9 F 28.7
Max 272.8 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.7 F 0.0
5:00 236.6 F 0.0
10:00 228.9 F 0.0
15:00 230.4 F 0.0
20:00 230.9 F 0.0
25:00 224.8 F 0.0
30:00 220.7 F 0.0
Drying Complete

FILLING: 1:06
HEATING: 20:49
STERILIZING: 4:00
VENTING: 2:58
DRYING: 30:00
TOTAL CYCLE: 00:58:53
CYCLE COMPLETE

03/01/2021 03:49 PM
ACCEPTED: X
REJECTED:
Operator: U01