
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 006642
8:28 20/11/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:24
PRESSURIZING 1:29
STERILIZING 2:37
136.1 C 220kPa 2:37
136.0 C 220kPa 3:07
136.4 C 222kPa 3:37
136.4 C 222kPa 4:07
136.3 C 221kPa 4:37
136.3 C 217kPa 5:07
136.3 C 223kPa 5:37
135.7 C 217kPa 6:07
135.7 C 217kPa 6:07
VENTING 6:07
AIR DRYING 7:02
CYCLE COMPLETE 7:28

Digital Signature #
15173DA84A701F09
