
Midmark M11 - v1.0.5
Total Cycles: 1109

Name: RITTER2
Sterilizer ID: V2313228
05/18/2022 12:14 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 124.9 F 0.0
2:00 127.2 F 0.0
4:00 141.3 F 2.3
6:00 180.2 F 6.6
8:00 205.0 F 11.7
10:00 216.9 F 6.1
12:00 230.6 F 11.3
14:00 230.7 F 8.1
16:00 241.6 F 11.9
18:00 244.1 F 12.8
20:00 256.3 F 19.3
22:00 267.4 F 26.2

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.0
1:00 272.6 F 29.9
2:00 271.9 F 29.2
3:00 271.1 F 29.1
4:00 272.6 F 29.9

Min 271.1 F 28.8
Max 272.7 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.4 F 0.0
5:00 220.3 F 0.0
10:00 226.4 F 0.0
15:00 243.6 F 0.0
20:00 236.3 F 0.0
25:00 222.4 F 0.0
30:00 220.1 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 23:28
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 01:01:26
CYCLE COMPLETE

05/18/2022 01:15 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____