
Midmark M11 - v1.0.5
Total Cycles: 308

Name: RITTER2
Sterilizer ID: V2313228
08/25/2021 08:00 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 148.4 F 0.0
2:00 147.7 F 0.0
4:00 167.4 F 0.0
6:00 193.3 F 0.0
8:00 209.8 F 0.7
10:00 221.8 F 3.7
12:00 232.7 F 7.8
14:00 243.2 F 12.2
16:00 239.8 F 10.3
18:00 237.7 F 9.0
20:00 240.3 F 10.1
22:00 249.8 F 15.2
24:00 261.4 F 21.7
26:00 271.9 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.4
1:00 272.5 F 28.7
2:00 272.6 F 29.2
3:00 271.8 F 28.3
4:00 272.9 F 29.1

Min 271.7 F 28.3
Max 272.9 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 172.0 F 0.0
5:00 224.6 F 0.0
10:00 194.6 F 0.0
Dry Cycle Aborted

FILLING: 0:55
HEATING: 26:31
STERILIZING: 4:00
VENTING: 2:57
DRYING: 11:43
TOTAL CYCLE: 00:46:06

08/25/2021 08:46 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____