
Midmark M11 - v1.0.5
Total Cycles: 1649

Name: RITTER2
Sterilizer ID: V2313228
10/25/2022 06:48 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 76.7 F 0.0
2:00 77.0 F 0.0
4:00 83.8 F 0.0
6:00 113.1 F 0.0
8:00 148.6 F 0.1
10:00 181.3 F 1.3
12:00 211.0 F 5.0
14:00 220.2 F 8.5
16:00 233.0 F 12.2
18:00 229.4 F 7.3
20:00 240.9 F 12.0
22:00 236.3 F 9.2
24:00 239.8 F 10.2
26:00 246.2 F 13.4
28:00 257.8 F 19.6
30:00 268.2 F 26.2

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.5
1:00 271.9 F 29.1
2:00 272.4 F 28.9
3:00 271.3 F 28.7
4:00 272.4 F 29.2

Min 271.1 F 28.3
Max 272.7 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 178.2 F 0.0
5:00 240.3 F 0.0
10:00 236.7 F 0.0
15:00 214.8 F 0.0
20:00 186.4 F 0.0
25:00 182.6 F 0.0
30:00 181.1 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 31:11
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 01:09:20
CYCLE COMPLETE

10/25/2022 07:57 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____