
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 001134
14:54 22/06/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:34
PRESSURIZING 2:44
STERILIZING 3:53
136.0 C 220kPa 3:53
136.4 C 225kPa 4:23
136.1 C 222kPa 4:53
136.3 C 224kPa 5:23
136.4 C 225kPa 5:53
136.1 C 223kPa 6:23
136.4 C 225kPa 6:53
136.1 C 223kPa 7:23
136.1 C 223kPa 7:23
VENTING 7:23
AIR DRYING 8:16
CYCLE COMPLETE 23:16

Digital Signature #
19DA305BE1BA52C6
