
Midmark M11 - v1.0.0
Total Cycles: 828

Name: RITTER 1
Sterilizer ID: V2178198
05/20/2021 08:45 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 163.4 F 0.0
2:00 164.8 F 2.0
4:00 174.5 F 5.1
6:00 195.6 F 9.2
8:00 213.4 F 6.6
10:00 224.9 F 7.4
12:00 236.6 F 12.0
14:00 235.6 F 8.9
16:00 241.6 F 10.7
18:00 246.3 F 13.1
20:00 258.0 F 19.4
22:00 268.6 F 26.1

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.1
1:00 272.6 F 28.8
2:00 271.6 F 28.0
3:00 272.6 F 28.9
4:00 272.2 F 28.4

Min 271.4 F 28.0
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 214.3 F 0.0
5:00 208.2 F 0.0
10:00 207.5 F 0.0
15:00 220.0 F 0.0
20:00 209.6 F 0.0
25:00 216.4 F 0.0
30:00 227.2 F 0.0
Drying Complete

FILLING: 1:08
HEATING: 23:13
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 01:01:39
CYCLE COMPLETE

05/20/2021 09:46 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____