
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001354
11:24 18/11/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:27
PRESSURIZING 4:27
STERILIZING 5:35
136.0 C 221kPa 5:35
136.2 C 223kPa 6:05
136.1 C 223kPa 6:35
136.3 C 224kPa 7:05
136.2 C 223kPa 7:35
136.4 C 223kPa 8:05
136.4 C 225kPa 8:35
136.2 C 223kPa 9:05
136.2 C 223kPa 9:05
VENTING 9:05
AIR DRYING 10:00
CYCLE COMPLETE 20:10

Digital Signature #
8606524FBEACD6E2
