
Midmark M11 - v1.0.5
Total Cycles: 1436

Name: RITTER2
Sterilizer ID: V2313228
08/30/2022 03:38 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 133.6 F 0.0
2:00 134.5 F 0.0
4:00 145.7 F 1.3
6:00 175.1 F 4.9
8:00 199.1 F 9.5
10:00 223.0 F 8.1
12:00 231.5 F 9.6
14:00 236.0 F 9.7
16:00 242.1 F 12.7
18:00 241.7 F 11.8
20:00 255.3 F 18.7
22:00 266.9 F 26.0

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.0
1:00 272.0 F 29.3
2:00 272.4 F 29.7
3:00 272.0 F 29.0
4:00 272.6 F 29.7

Min 271.5 F 28.8
Max 272.9 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 178.0 F 0.0
5:00 214.3 F 0.0
10:00 225.2 F 0.0
15:00 229.5 F 0.0
20:00 232.6 F 0.0
25:00 223.9 F 0.0
30:00 223.4 F 0.0
Drying Complete

FILLING: 1:28
HEATING: 23:23
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 01:01:58
CYCLE COMPLETE

08/30/2022 04:39 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____