
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
2.5uS / 1.6ppm
CYCLE NUMBER 003682
11:14 25/05/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:09
PRESSURIZING 2:12
STERILIZING 3:13
136.4 C 223kPa 3:13
136.2 C 222kPa 3:43
136.4 C 225kPa 4:13
136.4 C 223kPa 4:43
136.2 C 221kPa 5:13
135.9 C 220kPa 5:43
136.4 C 219kPa 6:13
136.2 C 223kPa 6:43
136.2 C 223kPa 6:43
VENTING 6:43
AIR DRYING 7:35
CYCLE COMPLETE 17:42

Digital Signature #
D1492121C576CA70
