
Midmark M11 - v1.0.0
Total Cycles: 4078

Name: RITTER 1
Sterilizer ID: V2178198
12/04/2023 12:13 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 152.5 F 0.0
2:00 150.0 F 2.0
4:00 158.0 F 5.1
6:00 189.8 F 8.3
8:00 199.8 F 13.0
10:00 215.5 F 5.4
12:00 226.0 F 9.7
14:00 234.7 F 9.6
16:00 238.4 F 10.9
18:00 236.9 F 8.5
20:00 248.4 F 14.6
22:00 259.3 F 20.7
24:00 269.7 F 27.3

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.5
1:00 272.3 F 29.3
2:00 272.5 F 29.0
3:00 271.5 F 28.5
4:00 271.2 F 28.6

Min 271.1 F 28.4
Max 272.6 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.3 F 0.0
5:00 221.6 F 0.0
10:00 229.2 F 0.0
15:00 229.2 F 0.0
20:00 227.0 F 0.0
25:00 228.7 F 0.0
30:00 229.2 F 0.0
Drying Complete

FILLING: 1:03
HEATING: 25:04
STERILIZING: 4:00
VENTING: 3:44
DRYING: 30:00
TOTAL CYCLE: 01:03:51
CYCLE COMPLETE

12/04/2023 01:16 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____